

# **METHOD AND APPARATUS FOR NON-CONTACT MONITORING OF CELLULAR BIOACTIVITY**

## **BACKGROUND OF THE INVENTION**

The present invention pertains to a method for non-contact (non-  
5 invasively) monitoring of cellular bioactivities by three means:

- a)- Monitoring impedance variations which occur due to the  
transition movements of ionized molecules across the cells  
membrane.
- 10 b)- Monitoring impedance variations which occur due to  
differential concentrations of ionized molecules at both sides of  
the cells membranes; this is the basis of cellular impedance.
- c)- Monitoring impedance variation which arise from cellular  
movements or flow within the body.

Monitoring micro-movement activities of concealed, non-metallic objects  
15 in free atmosphere, is extremely difficult with previously applied  
technologies.

Current motion detectors are classified in two general categories:

First is the active type – a technology based on transmitting a signal of  
ultrasound, laser, or electromagnetic waves (EMW) at the target of  
20 interest, and comparing the transmitted signal with signal reflected from  
the target to detect movements (eg. Radar).

Second is the passive type – Such as, infrared motion detector used in  
access control, is based on capturing target activities such as the heat  
changes.

The aforementioned techniques have limited sensitivity due to the need to transmit the detection signal through the ambient atmosphere, where ambient noise and interference may distort the signal and thereby compromise the accuracy of the detection system.

- 5 The creation of a closed system that can monitor the movements activity of concealed bodies and objects has been achieved by means of exploiting the transmission line impedance match and mismatch phenomena. The patent to Haj-Yousef (US Patent 6359597), has demonstrated the method that can be used to achieve a closed system motion detector. The prototype which
- 10 was used in support of the said Haj-Yousef Patent reached a sensitivity of about 100 parts per million (PPM), therefore it was intended to monitor the relatively large object movements.

Sensitivity has been estimated by the ratio of the impedance variation ( $\Delta Z$ ), which occurs due to the target movements relative to the total

15 impedance ( $Z$ ) of the surrounding media, which contains that target.

$$\text{Sensitivity (S)} = \Delta Z / Z$$

A subsequent prototype was completed which demonstrated a sensitivity of about 0.1 PPM. This means that for 50-ohm load impedance, a 5 micro-ohm impedance variation has been traced.

- 20 Laboratory observation with subsequent prototypes indicate that sensing micron and even the nano-scale movements of objects is possible. Work to date with this technology indicates the potential for achieving sensitivity in the range of a few parts per trillion (PPT).

- 25 A number of un-expected results have been observed in this work, these include (by way of example and not limitation), monitoring the movements of hidden isolators (Glass, Plastic, sponge...etc.) through a non-metallic or partially metallic barrier; a similar sensitivity is observed when monitoring

movements of metallic objects. Because the isolators are negatively effecting the resultant impedance of the surrounding (inspected) media when the isolators occupies part of it, which therefore reduces the components that characterizes the resultant impedance of the said media.

5 Sensitivity is not affected by normal or artificial airflow even if the scrutinized media is the ambient air. While the air moves, the surrounding air immediately occupies the same place. Therefore the airflow doesn't produce any variations in the resultant impedance.

10 Additionally, sensitivity is not effected by the presence flame within the inspected area.

From the aforementioned observations, additional important applications have been identified.

## **SUMMARY OF THE INVENTION**

15

When stable high frequency (HF) electromagnetic waves (EMW) travel outwardly along a transmission line (coaxial cable, dual strip-lines...etc) to a balanced antenna which surrounds (adjacent) the scrutinized media that contains the target of interest, a specific power value of the EMW will be released from the transmission line and completely absorbed by the load (media being monitored) due to the impedance match level between the load and the EMW source. Any impedance mismatch results in a different specific power value which will not be released from the transmission line, rather, when the mismatch waves reaches the end of the transmission line, it is reflected back toward the EMW source.

20

25

It is well known that the maximum power absorbed by the load occurs when the load impedance and the EMW source impedance are equal (fully matched). Therefore monitoring these two power values provides an

opportunity to obtain values relating to the actual degree of match / mismatch.

The first power value, which has been completely absorbed by the load due to the impedance match, is called the forward (incident) value and it can be sampled before being released from the transmission line. The second power value, which is reflected in phase and not released from the transmission line due to the impedance mismatch, is called the reflected value. These two bi-directional power values are generated only inside the transmission line, and they manifest on the basis of the transmission line impedance match and mismatch phenomena. These two power values can be sampled instantly and precisely by passing the transmitted EMW through a bi-directional coupler, which is connected in series within the said transmission line.

A matching network is used to buffer and tune the load impedance to approximately 50-Ohm ( $\Omega$ ), which is equal to the system impedance (transmission line and EMW source). The bi-directional coupler samples the instant power values of the forward and reflected waves in voltage form, such as forward voltage  $V_F$ , and the reflected voltage  $V_R$ . These two voltage values are totally free from any ambient electro-magnetic interference (EMI) or noise, because the forward power value is sampled before being released from the transmission line, and the reflected power value is never released from the said line. Any slight movements within the inspected media by the target of interest will vary the resultant characteristic impedance of the said media, which will also vary positively or negatively the degree of impedance match and mismatch. By using two DC (direct current) blocking capacitors functioning as a high pass filter (HPF) connected to the rectified outputs of the bi-directional coupler, the variable components of the  $V_F$  and  $V_R$  voltages, which contain the useful

indication about the target movements (activities), are only crossing these capacitors toward the next processing stage. The variable components of the  $V_F$  and  $V_R$  voltages have a symmetric non-proportional relationship, which means that when the  $V_F$  signal increases, the  $V_R$  signal decreases and vice versa, and a combined differential signal occurs. By directing these two extracted variable voltages to the inputs of the Differential amplifier (DA) or to what is so called “Instrumentation amplifier” (IA); an indication of the target movements and activities can be obtained and a close type monitoring system is established.

By way of example and not of limitation, the method of the present invention comprises the usage in the following applications:

1. Non-contact monitoring of bodily Hemodynamics:  
The Hemodynamics of the blood flow within the capillaries or main vessels (veins / artery) contains important data which reflects mechanical activities of the vital organs. Since the heart and the lung are mechanical organs, a way to detect their mechanical performance is vital and will become an essential diagnostic tool.  
The principal present-day methods used to monitor vital hemodynamic activities within the body are:
  - a)- Heart rate sensors (Photo-Plethysmography - PPG): this is used to measure the cardiovascular pulse wave that's found throughout the human body. The pulse wave results in a change in the volume of arterial blood with each pulse beat. This change in blood volume can be detected in peripheral body parts such as fingertips or ear lobes. The technique consists of an infrared Light Emitting Diode (LED), which illuminates the tissue and Light Sensitive Detector (LSD), which has been tuned to the same color wavelength as the LED, and therefore it detects the amount of

light absorbed by the tissue. The beats per minute are calculated by timing the width of a pulse and scaling up to a rate of beats per minute.

5       b)- Pulse Oximetry: The principle of pulse Oximetry is based on the red and infrared light absorption characteristics of the oxygenated and the deoxygenated hemoglobin. Oxygenated hemoglobin absorbs more infrared light and allows more red light to pass through. This method uses two LED's, red and infrared instead of one infrared LED as in PPG, in a similar way it monitors the light absorption of blood within the fingertip or the ear lobe to acquire an indication of respiration activity.

10       c)- Ultrasonic blood flow Doppler: a transducer probe is used to beam ultrasonic waves into a specific vessel. The beam is reflected with slight frequency changes that are due to the speed variations of the blood flow; by tracking these frequency variations, the speed (non-directly the volume) of blood flow can be monitored.

15       Optical methods used for blood flow measurements have many limitations and disadvantages; for example, to obtain a suitable reading, in most cases it is necessary to warm the hands by rubbing to increase the blood flow. The following physical factors and diseases also disturb optical readings: Malpositioned Sensor (Penumbra Effect), Light Interference, Sensor site  
20       temperature, Fingernail Polish, skin color, Motion, Burns, Venous Motion, Venous Pulsations, Venous Congestion, Sickle Cell Anemia, Pressure Necrosis, Fetal Hemoglobin, Intravascular Dyes, Bilirubin, Low Perfusion, Localized Hypoxemia, Carboxyhemoglobin, Methemoglobin, Low oxygen saturation (SaO<sub>2</sub> less than 70%), Magnetic Resonance Imaging (MRI),  
25       Electrocautery.

Optical method used to detect the absorption variations of specific color, cannot detect both veinal and arterial blood at the same symmetrical

sensitivity, as the color of the oxygenated and deoxygenated blood is highly variable from person to person.

The following are the disadvantages and the limitations of the Doppler blood flow technique:

- 5    The Doppler transducer should be well secured to the vessel to be inspected; this disturbs normal blood flow characteristics. Moreover portable Doppler equipment can be only used to monitor the visible, superficial vessels. Doppler is still not able to monitor capillary blood flow such as fingertip or the ear lobe. Similar to optical method limitations,
- 10   Doppler cannot recognize blood flow variations of less than 1% over the full-scale bandwidth. The big physical size of the Doppler transducer is an additional disadvantage, specially for long-term monitoring.

At present, the traditional electro-cardio graph (ECG) outlines only five curves / three peaks (QRS, T and P). Diagnosing the ECG abnormality can

15   be obtained by monitoring changes in the observed signals.

The ECG is now used to identify a limited number of heart problems such as arrhythmia. The possibility of spurious results exist as in cases where the ECG show normal heart activity while the heart is completely dead, as in the case of electro-mechanical cardiac dissociation.

- 20   The present invention has overcome limitations associated with previous monitoring techniques. It is not influenced by any kind of ambient EMI or noise. Blood flow variations in the range of few parts per billion (PPB) can be monitored. Additionally, diseases which affect blood contents or intensity, do not bias the results.

25

The subject invention can provide non-invasive monitoring by fastening a small insulated probe at any point on the body and may even be placed above some layers of clothing or shoe.

This technology can be used to monitor blood flow waveform within any vein or artery to determine clotted vessels or arteriosclerosis. Positioning the sensor over an area of skin that contains the capillaries allows provision of extensive data concerning heart/Hemodynamic activities. The capillaries are the locus where the veins and the artery are coupled. Therefore this area contains signals which can be monitored to obtain information regarding heart and circulatory system function.

The sensitivity of the current prototype has reached a level that can trace the activities for any of the heart components such as the valves or the cavities contractions within the same magnified graph. At the same time, heartbeat rate and respiration activity can also be precisely monitored, since the inspiration causes thoracic pressure to decrease inside vessels; this slight pressure decrease is not available by other non-contact techniques.

Over few millimeters of skin at rest, a blood flow variations as low as few nano-liter can be traced by the subject technology.

The subject invention will provided a robust, reliable, and easily applied analytical tool in the field of medical diagnosis.

## 2. Non-contact monitoring nervous system bioactivities:

This invention (basic sensing technology) may be beneficially applied to a variety of applications in a broad spectrum of industries.

Before explaining the value of this conclusion, it is important to explain how this been deduced.

While testing the slight voluntary movements of the finger, by fully relaxing the palm above a 5 centimeter thick granite tile, and the motion sensor was placed beneath the said tile. It has been observed that the system is distinguishing the initial movement action before it is perceived by the subject. Apparently, the finger starts the movement. This infers that



the sensory/perceptual threshold is higher than the muscle threshold; this indicates that muscle movement occurs without conscious perception of the brain, rather activity precipitating movement, appears in the peripheral nervous system. When the sensor is positioned posterior region of the neck, a few millimeters from the Medulla Oblongata at the top of the Spinal Cord, the system has monitored many signals synchronized with neural/ sensory or motor activities such as the slight movements of the toe. Such observation indicate that the system is capable to monitoring more than simple muscle movement.

When head is positioned between two insulated sensor electrodes as depicted in FIG. 4, and the electrodes are distanced from the scalp by one-centimeter thick sponge, the sensors report brain signals which appear to reflect instant brain response to visual stimuli presented to the subject. The resulting graph appears similar to the standard signals obtained by classical EEG; eg. delta waves.

The subject sensing system has demonstrated, under different circumstances, that it is unaffected by any electric or magnetic interference or noise.

Additionally, the detected brain signal has been instantly synchronized with the observed brain activity sensed by conventional methods, thereby eliminating the possibility that the subject sensing technique is monitoring only blood flow variations as a result of the brain activity, functional MRI has shown that the pattern excitation results in maximum blood flow in the brain with a time delay of approximately 6 seconds.

From the preceding examples and observations this system has demonstrated the capability of sensing neural communications by monitoring molecular activity which occurs due to the chemical reactions within the neurons.

In the mechanics of the central nervous system(CNS), the communications between the brain and sensory or the motor neurons, is achieved by creating a chemical reactions within the neuron membrane where ionized potassium, sodium and chloride molecules are moving on both sides of the neuron membrane; these molecules create differential concentrations of polarized ions inside and outside the neuron which result in the synapse firing. This molecular movement is repeated many times along the nerve fibers until the massage reaches its target.

A collection of living cells always has properties of resistance, displacement capacitance, and impedance. When the cell is stable or at rest, there is a 70mV potential between inside and outside of the cell, potassium ions are concentrated inside the cell and sodium and chloride ions are concentrated outside the cell. The cell at rest also has an electric resistance of about 10k $\Omega$ /cm, and it has about 1k $\Omega$ /cm at action. Therefore the cell bioactivity produces three effects: molecular movement, impedance variation, and electrical potential.

Impedance Cardiography (Impedance Plethysmography), also known as thoracic electrical bio-impedance, is a method that has been used to measure superficial impedance changes in order to monitor internal bioactivity. Impedance measurement is achieved by introducing an electric current into the body surface and then measuring the corresponding voltage. The ratio of voltage to current gives impedance (ohms law). Any change in the region's conductivity produces a change in the resultant impedance, which is proportional to the amount of current flowing in that region. Separate electrode pairs for introduction of current and measurement of voltage are used; the outer electrode pair is used to introduce the current, the voltage is measured across the inner electrode pair. This method employs direct electrical contact with the patient. At

present the sensitivity is limited to about 0.1 to 0.01 ohm, therefore this method has been applied in monitoring large physiological activities.

A key feature of the present invention is the elimination of direct electrical contact with the patient; the subject sensing technology invention, facilitates acquisition and interpretation of micro and nano-ohm variations in load impedance via a closed monitoring system.

In view of the previously reviewed technical functionality, potential applications of the subject sensing technology invention, would not be limited by, but would include the following:

- 10 a)- Non-invasive encephalograph: which monitors brain bio-activities by tracking impedance variations and following molecular movement inside the brain. Monitoring brain bio-activities can be used in non-medical sectors also, such as criminal investigations; by so called “Brain fingerprinting”, a method created by Dr. Lawrence Farwell to identify the perpetrator of a crime, by associating physical evidence from the crime scene with the evidence stored in the brain, and measuring brain wave responses to crime-relevant words or pictures presented to the suspect.
- 15
- 20
- 25 b)- Non-invasively monitoring nervous system bio-activity: this can help in the diagnosis of nervous system or muscle diseases. This subject sensing technology may be of significant value in realizing intelligent prosthetic limbs and sensory organs by acting as the interface for direct neural activation of prosthetic devices.

Functional magnetic resonance imaging (f-MRI) is also a method used to study the blood-flow volume inside the brain and thereby, indirectly, brain activity. By taking many short interval anatomic pictures of the brain, blood-flow to different regions can be observed as changes in the sizes of blood vessels. The assumption made here is that the areas of the brain, which are in use, will require more blood, and if they are using more blood the blood vessels will be larger. Researchers look at changing sizes of the blood vessels then infer that particular regions of the brain are being used at particular times.

Many methods are currently used to monitor the brain bio-activity, such as the Electro-encephalograph (EEG), which employs a dozens of bulky electrodes being attached to the scalp through salted gluing gel; this technique is used for capturing the brain bioelectricity. Due to the ambient EMI and noise, the EEG monitors only brain signals that are larger than 1 micro-volt can be distinguished.

Most of the current methods used to monitor cellular bio-activity are focused on direct capture of the cellular bio-electricity; as an example, Electro-cardiography (ECG), Electro-encephalography (EEG), Electro-myography (EMG), Electro-nervography (ENG), Electro-gastrography (EGG)...etc.

Tracking electrical signals which occur in a specific combination of cells (tissue) has many limitations which degrade the value of the extracted data. For example, different sources of bio-electricity are crossing the same frequency bandwidths, as in the EMG and ENG. Narrowing the monitored bandwidth is one possible solution, but it omits much relevant data. Also, the level of sensitivity is impacted by the level of the ambient EMI and noise, which is involved in achieving good signal to noise ratio (SNR). Additionally, to monitor cellular bio-electricity, a direct electrical contact

with the patient is currently required. Therefore, additional precautions for the patient safety are vital to avoid any threat to the patient from an accidental system breakdown which could result in an electrical shock as current monitoring systems are usually powered from the main AC power line. Attaching contact electrodes to the patient skin results in poor stability and introduces extraneous noise. The following are limitation factors: skin preparation (hair shaving), air bubbles within the conducting gel, electrode and lead motion artifacts, current leakage, electrode polarization specially in long term monitoring, large skin to electrode impedance, perspiration, EMI, Electrocautery, MRI...etc.

The presented invention provides a tool for non-contact monitoring of bio-activities of the central or peripheral nervous system (brain, spinal cord and spinal nerves...etc), the very low sensing threshold can track even very low power brain wave activity (brain whispers).

This method overcomes many limitations of currently deployed technologies by monitoring directly and non-reactively extremely low biological effects of microscopic particles within living cells.

### 3. Fetal Cardiography:

Potential applications include fetal-cardiography monitoring the fetal heart rate and the maternal contractions by tracking vital bioactivities of fetal organs.

### 4. Insect cardio, respiration, and general activity graph:

All types of insects conduct respiration and have blood circulation. Monitoring such microscopic movements is of significant research interest, but extremely difficult to achieve. The presented invention has the capability of monitoring this activity while the insect is allowed free movement inside a ventilated chamber under no stress of direct sensor contact.

This can be achieved by placing the anesthetized insect above the insulated sensing plate for establishing the insects reference baseline, subsequently a spectrum analyzer is used to evaluate the normal Insect bioactivity by determining its frequency bandwidths. Then, while the insect in its normal activities [applying Fourier theorem using band-pass filters (BPF) via digital signal processing (DSP)] the target frequencies bandwidth, are tracked and extracted from the signal complex which contains all of the insect signal artifacts including the bioactivities.

Monitoring the insect's vital signs within any recommended ambient condition is vital and will expand our knowledge of insect biology. The subject technology will provide an essential development tool in the field of pesticides technology.

## **BRIEF DESCRIPTION OF THE DRAWINGS**

15

Fig. 1 is a block diagram of the preferred embodiment

Fig. 2 illustrates an embodiment of the invention for monitoring the superficial hemodynamics through the finger; likewise monitoring can be obtained through any other region of the body surface.

20 Fig. 3 illustrates a cross-sectional view of the finger being monitored. The highest EMW intensity is found inside the finger and near the surrounding electrodes. The sensitivity grading lines (regions) also illustrated.

Fig. 4 illustrates an embodiment of the invention for monitoring the brain bioactivities (a single channel is shown); the probe assembly looks like the normal headphone.

25

Fig. 5 illustrates an embodiment of the invention for monitoring the bioactivities of the central and peripheral neural system (spinal cord and spinal nerves...etc).

Fig. 6 illustrates an embodiment of the invention for monitoring the vital signs of the fetal organs, such as the heart, the lung, and the maternal contractions.

Fig. 7 illustrates the preferred shape of the overlapped transmitting antenna that can be used to monitor the insect's bioactivity. The electrodes can be  
5 constructed from a copper clad fiberglass board (printed circuit board), wherein the 1.6 mm fiberglass can perform the required insulation.

Fig. 8 illustrates the assembly details of the preferred adjacent-type transmitting antenna 48 (sensor / probe).

10 Fig. 9 illustrates the internal details of the preferred ultra-narrow band pass filter, which comprises many parallel crystal ladder filter modules 34.

Fig. 10 illustrates an embodiment of the invention as a multi-channel monitoring of bioactivities.

Fig. 11 illustrates the preferred embodiment of the invention that maintains  
15 the high sensitivity while utilizing a low transmitting power of EMW.

Fig. 12 illustrates the preferred embodiment of the invention by adding non-directional coupler within the transmission line to produce a negative reference that reflects the instability of the produced HF EMW.

Fig. 13 illustrates the non-proportional characteristic of  $V_F$  and  $V_R$  over the  
20 impedance bandwidth of a tuned load. The sensitivity becomes maximum where the curves turn into exponentially sharp ( $V_{SWR} = 1$  to 1.5).

FIG. 14 illustrates samples of the output graphs being obtained by the present invention, with the exception of graph (a). The remaining graphs are genuine, absolutely raw, and have not been processed by any means;  
25 the graphs were obtained in sequence using a single channel analogue to digital converter (ADC), then they were manually added to the said figure; the illustrated graphs were obtained by placing the sensing probe at various points of the body of a 39 years old man. The graphs were extracted using

different frequency bandwidths, but more clinical research is still needed to define the best frequency bandwidths:

- 5           a) Represent standard ECG chart, and it is just shown here for the ease of comparison with the obtained graphs. Normally the raw ECG signal contains EMI, noises, and artifacts in addition to the useful signal. Therefore many signal processing is required to achieve an acceptable graph like the one shown here.
- 10          b) Represent the hemodynamic cardiograph obtained from the wrist, above the capillary vessels and through clothes, the equivalent QRS, T, and P peaks are obvious, they vary slightly from the ECG by the time bandwidth, since the ECG outlines heart's depolarization instead of monitoring heart's mechanical activity as in the presented invention. The  
15           corresponding bandwidth applied is from 1 to 25 Hz.
- c) Represent the hemodynamic cardiograph obtained from the wrist, above the main vessels (artery and vein), a very tiny rhythmic curves are clear at the top of the systolic peak. The corresponding bandwidth applied is from 1 to 25 Hz.
- 20          d) Represent the hemodynamic cardiograph obtained from the thigh, a very tiny rhythmic curves are clear on the right side of the systolic peak. The corresponding bandwidth applied is from 1 to 25 Hz.
- e) Represent the obtained heart beat cycle that can be used to  
25           calculate the heart beat rate, this is covered by the frequency bandwidth of 1 to 1.5 Hz.
- f) Represent the obtained respiration cycle that covers the frequency bandwidth of 0.2 to 0.3 Hz.



similar hemodynamic cardio and respiration graphs were obtained by putting a larger sensor (size of few centimeters) beneath a chair, through 10 to 15cm thick non-metallic barrier between the body of the person sitting on the chair and the sensor probe. Such way of remote hemodynamic monitoring can be very useful in few critical cases where it is vital to react immediately; for instance monitoring aged passengers in airplanes, or drivers of critical transportations such as pilots, astronauts, train drivers...etc. As well as monitoring patients in Intensive Care Units (ICU).

10

The numbers in the drawing are:

- 1 is a HF oscillator;
- 2 is a HF power amplifier;
- 3 is an ultra-narrow band pass filter;
- 15 4 is a transmission line;
- 5 is a rectifying diodes;
- 6 is a bi-directional coupler;
- 7 is a matching network;
- 8 is a transmitting cable;
- 20 9 is a Balun (balanced to unbalanced transformer);
- 10 is a transmitting electrodes (sensor probe);
- 11 is an inspected region;
- 12 and 13 are forward  $V_F$  and reflected  $V_R$  voltages, respectively;
- 14 and 15 are RF suppression chokes;
- 25 16 and 17 are DC blocking capacitors - HPF;
- 18 and 19 are the load resistors of the HPF;
- 20 and 21 are the extracted wavering (variable) signals of the  $V_F$  and  $V_R$  voltages;

- 22 is a differential or instrumentation amplifier;  
23 is a differential signal;  
24 is an analogue divider;  
25 are the divided outputs;  
5 26 are active filters;  
27 are output amplifiers;  
28 are output ports;  
29 is an electrical insulator;  
30 are fixing arms;  
10 31 are connecting wires;  
32 are overlapped transmitting electrodes  
33 is a multi-port HF power splitter;  
34 is a crystal ladder filter module;  
35 is a multi-port HF power combiner;  
15 36 is a bi-directional coupler with HF outputs;  
37 is an input filter (ceramic filter);  
38 is a HF selective amplifier;  
39 is a crystal ladder output filter;  
40 is a HF demodulator;  
20 41 is a HPF;  
42 is a HF forward power;  
43 is a HF reflected power;  
44 is a non-directional coupler;  
45 is a signal produced by the non-directional coupler;  
25 46 is a linear amplifier;  
47 is a negative reference;  
48 is a transmitting antenna (transducer / probe)

## DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENT

According to the present invention, the device can be described by referring to the drawings and more particularly to FIG. 1. The HF oscillator 1 is used to produce a fixed sinusoidal frequency, means to achieve a very stable and low noise EMW energy in the frequency range of 1 to 300 mega-hertz (MHz) with an output power of less than one milli-watt (mW). The produced EMW is then amplified to the desired power level ranging from 1 to 100 mW by the HF power amplifier 2. The ultra narrow band pass filter 3 is used to clean the produced EMW from noises.

The purified EMW then passes through a bi-directional coupler 6, which is connected in series within a transmission line 4 (coaxial cable, dual parallel wires, or strip-lines). The bi-directional coupler 6 is used for instant sampling both the internals forward and reflected power values, which are generated inside the said transmission line 4. The EMW then passes toward the matching network 7, which is used to tune and buffer the 50-ohm ( $\Omega$ ) impedance of the HF oscillator 1 with the characteristic impedance of the load 11 (the region of the body being monitored). Due to its structure, the matching network 7 furthermore will act as a harmonic reject filter, which can be built from any of the popular types L, PI, or T filter networks.

The released EMW from the matching network is then introduced to the said load directly by a balanced type-transmitting antenna 48, or it can be introduced to the load indirectly via transmitting coaxial cable 8 through the same antenna 48. The said antenna 48 as depicted in FIG. 8 consisted of pair of electrodes 10 made from insulated pieces of metallic sheets or

wires. The balanced antenna **48** is connected to the coaxial cable **8** by balun **9** (balanced to unbalanced transformer).

The geometry size of the electrodes **10** defines the preferred coverage area plus the desired sensitivity depth, wherein the larger electrodes (L and W) will cover more surface area, and the increase of distance (D) between both electrodes will increase the effective depth of the sensitivity within the media being monitored. For example: monitoring the superficial capillary blood flow within the finger as depicted in FIG.2, an electrode length (L) of about 5 millimeters (mm) by a width (W) of about 3 mm, having a distance (D) between both of the electrodes of about 2 to 3 mm, has seemed to be sufficient to obtain a satisfying results.

A dual HPF consisting of capacitor **16, 17** and resistor **18, 19** connected to both outputs **12, 13** of the bi-directional coupler **6**, the capacitors **16, 17** will only allow the variable (wavering) voltages to pass through, and the direct current (DC) will be rejected. The extracted variable voltages **20, 21** are imitating the impedance match and mismatch variations occurred in-between the load **11** and the HF oscillator **1**.

The extracted variable voltages **20, 21** are then combined together by a Differential amplifier (DA) **22** or to what is so called instrumentation amplifier (IA). Any ambient EMI could even reach the transmitted antenna **48**, affects both of the DA inputs **20, 21** evenly and by the same phase, therefore such external common noise will be highly rejected due to the high common mode rejection ratio (CMRR) available for this type amplifier, which now has been reached to more than 134 decibels (dB). Therefore the resulted combined signal has become to be so pure and unsusceptible to any external EMI or noise.

The output signal **23** thereafter directed to an analog divider **24**, which produces multi outputs **25** that each mirrors the same characteristics and

parameters of the input signal **23**. The analogue divider **24** in particular is required when the same signal contains many vital parameters and indications. For example the signal obtained by monitoring the hemodynamic activities of capillary vessels contain a lot of information about the mechanical heart activity beside the respiration cycle, therefore dividing the signal to two channels, each of them represents a specific activity that can be discriminated by limiting the expected frequency bandwidth. This is can be achieved by the subsequent use of an active filter **26**. The active filter **26** can be established by means of operational amplifiers with a few passive components such as resistors and capacitors, or by using of modern computerized technology, such as the digital signal processing, these circuits can achieve the low pass, high pass, band pass, or band reject filter.

However the produced signal still needs to be amplified to a sufficient level that can drive the next analytical circuits, this can be performed by the output amplifier **27**.

The general description explained above describes only the general functioning of the device. Utilizing standard readymade blocks, which are popular and widely available in the market cannot achieve sensitivity better than few parts per thousand only.

Consequently the system designer should take a few significant measures into consideration, which is so vital to achieve a system with an ultra high sensitivity.

The HF section (**1-10**) has become to be the most critical part that can define the final sensitivity. The design of HF oscillator requires more concernment about the noise floor. Manufacturers of the HF oscillators are now showing more attention for reducing the phase noise and for the enhancement of the long-term stability. Nevertheless as an example, an

ultra-low-noise RF oscillator with a noise floor of about -174dB has been established by the American Wenzel Inc. (ultra-blue low-noise oscillator series). Such oscillator is an excellent choice and exceeds the requirements, but it's output power of about 0.5 mw still very low.

- 5 The produced EMW power will need amplifying to become usable. This can be achieved by using of HF power amplifier. Using any of the popular hybrid wide band RF amplifier for this stage is useless. Such type of amplifier produces a lot of noises, which contaminate the amplified EMW. The use of narrow-band (selective), and very low-noise HF amplifier **2** is
- 10 vital to establish a high quality EMW. Nevertheless the produced EMW still need to be purified from the noises, which occurred internally by the amplifier and the oscillator circuits. Using of narrow band pass filter **3** will help, but the traditional LC (inductive and capacitive) resonant type will not achieve a satisfying quality, due to the limitation of there low quality
- 15 factor, which cannot exceed a few hundreds. A quality factor in millions can only be established by using of the crystal ladder filters (CLF) **34** as depicted in FIG.9. In general the present days crystals cannot tolerate driving powers above 10mW.

- In many applications it is required to use a higher power than the crystal
- 20 limit. Splitting the EMW energy by a HF power splitter **33** to a few matched and isolated ports will allow dividing the high EMW power to many matched and paralleled multi-order CLF **34**, thus by combining the filters outputs together by a HF power combiner **35**, a very clean and noise-free high power of EMW is produced.

- 25 Few many relationships should be maintained by the system designer, which is needed to achieve the highest sensitivity. For instance the produced DC voltage value of  $V_F$  **12** should be as large as possible, but to achieve a good differential symmetry between both voltage values of  $V_F$

12 and  $V_R$  13, the coupling coefficient of the bi-directional coupler 6 should be in the range of 30 to 40 dB.

The sensitivity threshold of the DA (IA) 22 is limited to a few nano-volts (nV) due to the DA/IA self noise. One of the best IA that has 1.6  
5 nV/(root)Hz self noise voltage which equals about 10nV RMS (Root Mean Square) and for the bandwidth of 0.1 to 100 Hz is INA 166 made by Texas Instruments Inc. Therefore the lowest input voltage which is required to achieve a good signal to noise ratio has to be larger than 100 nV, and this is ten times higher than the amplifier's internal noise.

10 The IA self-noise can be reduced many times by decreasing the circuit's temperature whenever such extremely high sensitivity is desired. This can be achieved by cooling the device through keeping the circuits inside a liquid nitrogen container. This way highly reduces the self-noise of the circuits by means of reducing the thermal, Johnson, and flicker noises.

15 If the sensitivity of about 1PPM is required, and the IA has 100nV minimum input voltage, therefore the bi-directional coupler should produce 0.1Volt DC value for the  $V_F$ .

$$V_F = 0.1V = 100nV \times 1,000,000$$

The 0.1V forward voltage embodies the 50-ohm ( $\Omega$ ) impedance of a  
20 matched load, consequently the 100nV variations (wavering) in the  $V_F$  voltage can be considered argumentatively to represent a 50 micro-ohm ( $\mu\Omega$ ) impedance variations in the same load.

$$50 \mu\Omega = 50\Omega / (0.1V/100nV) = 50\Omega / 1E6$$

This general way of assumption is not so accurate; since there is no such  
25 accurate tool available at the present days that could be relied on to measure such very low impedance variations.

The practical observations demonstrated that the actual impedance sensitivity for a matched load is much better than what have been

estimated above, because working at a good degree of impedance match of 1 to 1.5 VSWR (voltage standing wave ratio), where the curve for the  $V_F$  and  $V_R$  becomes exponentially sharp as depicted in FIG. 9, the slightest changes in the load impedance will lead to the highest changes for the  $V_F$  and the  $V_R$  values. The sensitivity decreases by increasing the degree of impedance match and vice versa.

Moreover increasing the DC-Voltage value of  $V_F$  increases the final system sensitivity. Due to the coupling coefficient limitation, the increase of EMW power being transmitted has been demonstrated to be the proper solution that increases the DC-Voltage of  $V_F$ .

For example if the IA has 100nV minimum input voltage, and if the DC value for the  $V_F$  equals 10V, therefore the resulted sensitivity will be:

$$\text{Sensitivity} = 10\text{V}/100\text{nV} = 10\text{PPB}.$$

Reducing the IA self-noise and enhancing the purity of the EMW can highly improves the final system sensitivity.

Actually in some cases it is not recommended to increase the transmitted power, for example in portable applications the power supply consumption is a very important factor, that's why the output power should be reduced. Also small insects cannot tolerate high RF powers while monitoring the their bioactivities. Also for a safety reason, according the regulations of the Federal Communications Commission / USA, the maximum permissible uncontrolled power exposure at 30 MHz for a period of 30 minutes, should not exceed 180mW/cm<sup>2</sup>.

Therefore acquiring a high DC Voltage value for the  $V_F$  from a low transmitted power is possible by using a bi-directional coupler that has un-rectified outputs as depicted in FIG.11, this means that the forward and reflected powers have to remain in there HF format without any demodulation. This enables the use of an ultra-narrow band (selective) RF



amplifier 38. While these tiny powers remain in the HF format, the amplifying is possible without the risk of adding an extra noise generated by the amplifier's circuit 38 to the amplified signal 42, 43. Narrowing the bandwidth of the amplifier highly reduces the amplifier self-noise, and therefore enhances the amplified signal purity.

In low-power applications, the signal being tracked is lesser than the amplifier self-noise, therefore by using an ultra-low noise and selective RF amplifier 38 along with many ultra-narrow BPF 37, 39, (CLF, LC, and ceramic filter), the purity of the amplified signal remains as the un-amplified one.

Rectifying (demodulating) the amplified HF powers can be achieved by using dual matched Schottky type diodes along with fining capacitors 40. Using P-type zero bias Schottky detector diodes is necessary for achieving a high rectifying linearity in a wide range of input voltages, and because of their own low-flicker noise.

It has been noticed that the system is susceptible to rough vibration artifacts, which therefore affects the mechanical stability of the HF oscillator circuit, this is because the center frequency of crystal oscillator 1 is very susceptible to mechanical vibrations. Moreover the matching network 7 consisting of frequency dependent components (inductors and capacitors), so any changes in the oscillator frequency, leads to instability (deviation) in the resultant impedance match. Therefore the weight of the oscillator circuit should be lightweight as much as possible, it has to be surrounded and fastened inside the device by placing it in sponge compartment that establishes a vibration absorber.

In order to prepare the system to be implemented in any application, a few tunings and modifications are required. In general the essential preparations are based on choosing the proper transmitting antenna 48,

which is used to introduce the EMW into the region of the body being monitored **11**. Likewise it is necessary to define the preferred sensitivity, as well as adjusting the frequency bandwidth to cover the expected bioactivities being monitored.

- 5 The final sensitivity can be easily tuned by adjusting the gain of the output amplifier **27**, and the frequency bandwidth can be tuned by adjusting the components of the active filter **26**, or by modifying the parameters of the DSP software.

However the shape of Electrodes **10** generally determines the type of the  
10 intended application. Each application requires different electrodes **10** with different size, shape, and insulation thickness.

Monitoring the tiny superficial blood flow fluctuations within the concealed capillary vessels requires more attention. The capillary blood flow within the superficial vessels (skin) at rest, has estimated to be about  
15 1 micro-liter per second ( $\mu\text{L/S}$ ) for each square centimeter, also the actual fluctuation in the capillary blood flow doesn't exceed 10% of the total volume flow. Consequently the blood fluctuates by about  $0.1 \mu\text{L/S}$ .

Moreover the hemodynamic cardiography monitors the instant capillary blood flow within the bandwidth of 0.1 to 100 Hz, therefore the upper  
20 frequency limit (100Hz), which represents the fast blood flow variations, outlines the 1 nano-liter (peak to peak) variations in the blood volume for each 10 milli-second ( $\text{nL}/10\text{mS}$ ).

The transmitting antenna (probe) **48** which is intended to monitor the superficial bioactivity such as the capillary blood flow within the skin,  
25 comprises of dual symmetrical electrodes **10** (FIG.8) made from thin sheet of metal, that have relatively a similar length (L) and width (W) of each electrode of about 2 to 5mm, and a distance (D) between both electrodes of a few milli-meters. The insulation layer **29** can be made from any thin

plastic or rubber sheet of less than 1mm thickness, means to achieve proper electrical isolation.

Monitoring a more deep bioactivity (brain, CNS, and fatal, FIG.4, 5, and 6) requires larger size electrodes 10 in the centimeter range, likewise an extended distance between the electrodes should be considered too, also a thicker insulation should be prepared.

The relationship between the bioactivity depth, electrode size, distance between electrodes, and the insulation thickness is a direct-proportional relationship. The purpose of using thick insulation is to reduce the sensitivity for the superficial bioactivity, and to increases the threshold sensation for a deeper bioactivity.

From the EMW propagation theory as depicted in FIG. 3 it is a well-known fact that EMW becomes attenuated by being away from the transmitting antenna, as well the direction of propagation turns to the surrounding objects that has the lowest impedance, wherein the surrounding objects luckily will act as a waves director. Therefore the EMW mainly propagates toward the nearest region of the body being monitored. Enlarging the electrode size will enlarge the inspected area being monitored. Likewise the distance enlargement between the electrodes enlarges the radius of the electromagnetic field being created.

The purpose of enlarging the insulator thickness is to keep electrodes away from the inspected region, to insure a deeper delivery of the EMW inside the body, and to reduce the high sensitivity (proximity effect) that occurs by positioning the electrodes very close to the body.

Moreover keeping electrodes away from that region which doesn't contain a large moving activity such as the skull, achieves another way of sensation to a deeper bioactivities that's directly affected by the impedance property of the cells, The cells (tissue) impedance varies from about

10k $\Omega$ /cm at rest, to about 1k $\Omega$ /cm at action. Therefore thickening the insulator **29** reduces the sensitivity to the tiny superficial bioactivities, and this is very important factor for monitoring the brain bioactivity without any significant interference with the natural blood flow within the skull.

5 The same arrangement can be done to monitor the CNS communications within the upper side of the spinal cord (Medulla Oblongata).

Placing the region being monitored between a two opposite transmitting electrodes as depicted in FIG. 3 insures the highest possible sensitivity, but such positioning reduces the impedance pre-matched flexibility from being

10 always ready for use in a different patient circumstances, this means that the distance between the opposite electrodes is defined by the thickness of the inspected region which varies between peoples, and therefore this varies the resultant load impedance, consequently any achieved impedance match cannot be valid for different positioning of electrodes. Therefore

15 this will force to employ an auto-tuning matching network, instead of permanent matching network **7**, which is applicable for many positioning circumstances. By employing the adjacent type transmitting electrodes as depicted in FIG. **8**, it is not required to retune the matching network **7** every time.

20 When the system has to be used out of clinic, by transmitting the patient cardiograms to a receiving unit that located in a hospital's emergency, also as in out-patient monitoring system which is used in the ambulance or the rescue helicopter, and the tele-patient monitoring (Bluetooth ® cardiography), or the soldier of future, were it is necessary to keep

25 watching the soldier health remotely within the battlefield. In such applications, which are running in shaking conditions that release many vibration artifacts, extra measures to maintain the monitoring stability are required.

The effect of vibration artifacts can be highly reduced by eliminating the output transmitting cable **8**, by means of connecting directly the sensing probe (antenna electrodes) **48** to the matching network **7**, this is to eliminate the vibration effect, which occurred due to the swinging in the transmitting cable **8**. Furthermore minimizing the device or at least the HF blocks (**1-10**) to a size that can be fit in the belt or the bracelet, which is fastened around the region of the body being monitored.

If it is needed to achieve multi-channel monitoring system as depicted in FIG. **10**, as for the multidirectional monitoring of the brain bioactivities, in a similar way of the traditional EEG. The produced power of EMW can be divided to many symmetrical ports that each port continues independently to all of the following stages. Individual CLF **34** is sufficient for each port, because the 10mW of EMW power is enough and sufficient for driving each port.

A dual-channel monitoring system is effective for subtracting the undesired signals, for example if the blood's flow affects the signal being obtained while monitoring the CNS bioactivities, an additional sensor can be used to monitor only the blood flow in that region of the body which doesn't contain any other activities, and then the resulted blood flow signal can be subtracted from the first signal being obtained by the CNS sensor. By this way, monitoring specific activity is possible even if it is founded in a region that contains undesired artifacts. Likewise it can help in monitoring the fetal activities without being influenced by the mother bioactivities.

Reducing the effect of internal noise and instability on the final sensitivity can be achieved by adding a non-directional coupler **44** within the same transmission line **4** as depicted in FIG. **12**. The output signal **45** of the non-directional coupler after being rectified is used to estimate the instability of

the produced EMW. The output signal **45** of the non-directional coupler **44** has no phase characteristics, and therefore it reflects only the amplitude instability of the produced EMW. By extracting the variable components through the use of DC blocking capacitor, the amplified signal **47** will  
5 contain the necessary data about the HF instability, and so it can be used as a negative reference in the final signal processing stage.

The very high sensitivity is required by a few limited applications, such as for monitoring the bioactivities of the very small insects like the ants. For a two-centimeter cockroach, a sensitivity of few PPM is sufficient to acquire  
10 good satisfying results.

At this time it is impossible to apply this technology for monitoring the bioactivity of an individual cell. This technology is still nascent, and it is intended now for monitoring the bioactivities for a large amount of cells that are combined in the tissue. In the near future and by the presented  
15 technology, the sensitivity could reach the capability of monitoring even the plant leaf bioactivities.